



Weqas Report Interpretation

Samantha Jones / Gareth Davies

Summary of Talk

- Weqas report format
 - *Target value
 - *Performance Criteria
- Basic statistics
 - *Imprecision
 - *Inaccuracy

Weqas Report Format


Quantitative Programmes

- Serum Chemistry
- Lipids
- Bilirubin
- ED Toxicology
- Urine Chemistry / Oxalate & Citrate
- Blood Gases / Co-oximetry
- Endocrine
- Haematinics
- Cardiac Marker
- BNP / NT Pro BNP
- HbA1c
- Homocysteine
- Bile Acids
- Serum ACE
- Serum hCG
- Porphyrin
- Ammonia
- CRP / POCT CRP
- TDM / IS
- POCT Creatinine
- Procalcitonin
- D-dimer

General Overview report of lab (section) SDI scores

Distribution Date: 24/10/17. Final. Report Issued: 28/11/17

This Distribution	
Overall Lab SDI:	0.24
Median All Laboratory:	0.38
97.5th centile:	2.27



All SDI Ranges	
< 1	Good
1 - 2	Acceptable
> 2	Poor

---- Median ● Lab SDI — 97.5th

Section SDI scores for this distribution

Section	Cobas e601
Overall	0.24
Cortisol	0.24
Progesterone	0.36
Oestradiol	0.06
Testosterone	0.49
Total T4	?
Total T3	?
Free T4	0.07
Free T3	0.15
TSH	0.32
LH	0.19
FSH	0.33
Prolactin	0.16

click for further details on individual analyte performance

Analyte detailed report

Scheme: Endocrine, Distribution Code: S245. Distribution Date: 24/10/17. Final Report Issued: 28/11/17							
Cortisol (nmol/L)	1	2	3	4	5	6	Analyte SDI
Reported Result	288.1	170.2	353.7	370.2	1064.0	72.0	
Method Corrected Result	288.10	170.20	353.70	370.20	1064.00	72.00	
Elecsys/E Module	Mean	306.62	184.24	381.54	396.64	1129.40	72.25
	SD	15.54	15.43	27.02	30.21	60.83	1.09
	Number	5	5	5	5	5	4
	Uncert.	6.950	6.903	12.083	13.512	27.204	0.545
Cobas E Module	Mean	306.62	184.24	381.54	396.64	1129.40	72.25
	SD	15.54	15.43	27.02	30.21	60.83	1.09
	Number	5	5	5	5	5	4
	Uncert.	6.950	6.903	12.083	13.512	27.204	0.545
Overall	Mean	282.70	163.62	347.38	359.41	1083.57	69.54
	SD	22.83	12.82	22.16	33.15	50.95	3.90
	Number	19	18	18	19	18	10
	Uncert.	5.237	3.022	5.224	7.606	12.009	1.233
Reference Values ID-LC-MS/MS	298.36	177.20	356.88	377.36	1065.45		
Ref. Value Uncertainty	6.100	3.600	7.300	7.700	21.800		
Non-scoring Reference Values							
WeQas SD	30.30	18.19	36.87	38.10	111.75	8.62	
SDI	0.18	0.36	0.17	0.28	-0.18	0.29	0.24

Ref Values used for SDI and Sigma Scoring

Performance Criteria


Total Error

SDI is a measurement of your total error and will include both inaccuracy and imprecision.

This Distribution S245
Your average analyte SDI for the 6 samples is 0.24

Previous SDI

Distribution S245




.... Median — Your SDI — 97.5th

Please note: Linear regression uses CF corrected data.


SDI Score

Analyte detailed report

This Distribution S245



Previous Distributions



Precision

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Sy.x = 6.13 nmol/L	Sy.x	4.46	27.15				
IS = 1	IS	2	53	0	0	0	0

Sy.x is the average deviation from the best fit line and is an index of scatter.

Precision Key

IS score	Interpretation
0 to 10	Good
11 to 150	Acceptable to Warning level
> 150	Unacceptable (including Curvilinear Data)

Accuracy

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Systematic proportional error (calibration) -2.61%	Proportional (%)	2.11	3.08				
Systematic constant error (blank) 12.04 nmol/L	Constant (nmol/L)	- 14.46	- 5.24	0.00	0.00	0.00	0.00

Bias includes components of proportional and constant errors. A proportional bias suggests an error of calibration whilst a constant bias suggests a blank error. Mixed errors will include significant components of both.

Current slope and intercept
 $y = 0.97x + 12.04$

Previous slope and intercept values




How do we set Performance Specifications and Target Values?

The Weqas Report

Target values used in Statistical Analysis

Hierarchy

- 
- Reference values – used for bias plot /SDI calculation and σ score
 - Method mean – used for SDI calc if no ref and $n > 8$
 - Overall mean – used for SDI calc and bias plot if no ref and $n < 8$
 - Analysers mean – on report for information only

Generation of Reference Target Values

- Specialist Laboratories Required
- Accredited to ISO17025 and ISO15195
- Limited number of laboratories worldwide


Value of Reference Targets

- Traceable to higher order
- Establishes method traceability for the lab – requirement of ISO 15189
- Highlights the pitfalls of using the trimmed overall mean as an accuracy target in EQA Schemes
- Useful in the post market vigilance of the IVD - Directive
- Required for UK MAPS


Determining analytical quality specifications

Allowable Total Analytical Error (TEa) encompasses both imprecision and inaccuracy.

Analytical goal Hierarchy




Interlaboratory variation FT4




Relationship of analyte conc. to SD follows a polynomial equation for most analytes and is non linear therefore using fixed % TE is not appropriate

“State of the art” v Biology




Biological goals not achievable. Performance has not improved over time.

“State of the art” v Biology



Biological goals not achievable. Although
Performance has improved over time.

“State of the art” v Biology



Performance criteria = Target value +/- TE

Allowable TE = 2*Weqas SD

Quality Indicators

SDI = (lab result-target value)/WEQAS SD

To be within performance criteria SDI < 2

Sigma score based on UK MAPS criteria

Sigma metric: $[(TE_a - bias_{obs})/s_{obs}]$

Total Allowable
Error as defined
by MAPS

Laboratory bias
at critical level

Within run CV
(%)

Analyte detailed report

Scheme: Endocrine, Distribution Code: S245. Distribution Date: 24/10/17. Final Report Issued: 28/11/17							
Cortisol (nmol/L)	1	2	3	4	5	6	Analyte SDI
Reported Result	288.1	170.2	353.7	370.2	1064.0	72.0	
Method Corrected Result	288.10	170.20	353.70	370.20	1064.00	72.00	
Eleysys/E Module	Mean	306.62	184.24	381.54	396.64	1129.40	72.25
	SD	15.54	15.43	27.02	30.21	60.83	1.09
	Number	5	5	5	5	5	4
	Uncert.	6.950	6.903	12.083	13.512	27.204	0.545
Cobas E Module	Mean	306.62	184.24	381.54	396.64	1129.40	72.25
	SD	15.54	15.43	27.02	30.21	60.83	1.09
	Number	5	5	5	5	5	4
	Uncert.	6.950	6.903	12.083	13.512	27.204	0.545
Overall	Mean	282.70	163.62	347.38	359.41	1083.57	69.54
	SD	22.83	12.82	22.16	33.15	50.95	3.90
	Number	19	18	18	19	18	10
	Uncert.	5.237	3.022	5.224	7.606	12.009	1.233
Reference Values ID-LC-MS/MS	298.36	177.20	356.88	377.36	1065.45		
Ref. Value Uncertainty	6.100	3.600	7.300	7.700	21.800		
Non-scoring Reference Values							
WeQas SD	30.30	18.19	36.87	38.10	111.75	8.62	
SDI	0.18	0.36	0.17	0.28	-0.18	0.29	0.24

Please note: Linear regression uses CF corrected data.

Ref Values used for SDI and Sigma Scoring

Performance Criteria Target


Total Error

SDI is a measurement of your total error and will include both inaccuracy and imprecision.

This Distribution S245
Your average analyte SDI for the 6 samples is 0.24

Previous SDI

Distribution S245



SDI Score

The WEQAS Report

Statistical Indices

- Imprecision: $S_{y.x}$, r , IS
- Inaccuracy: Bias plot, $y=mx+c$

Interpretation of Imprecision Markers

Sy.x is the deviation around the best fit line and is an indication of scatter. It is used as an index of within run precision. This is given in the units of the analyte in question.

Correlation coefficient (**r value**) is also an index of within run precision. The closer to 1.0 the value is, the better the precision (less scatter about the best fit line).

IS score is derived from the correlation coefficient.


$$IS = (1-r)*10,000$$

Precision Key


IS score	Interpretation
0 to 10	Good
11 to 150	Acceptable to Warning level
> 150	Unacceptable (including Curvilinear Data)

Analyte detailed report

This Distribution S245



Previous Distributions



Precision

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Sy.x = 6.13 nmol/L	Sy.x	4.46	27.15				
IS = 1	IS	2	53	0	0	0	0

Sy.x is the average deviation from the best fit line and is an index of scatter.

Precision Key

IS score	Interpretation
0 to 10	Good
11 to 150	Acceptable to Warning level
> 150	Unacceptable (including Curvilinear Data)

Accuracy

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Systematic proportional error (calibration) -2.61%	Proportional (%)	2.11	3.08				
Systematic constant error (blank) 12.04 nmol/L	Constant (nmol/L)	- 14.46	- 5.24	0.00	0.00	0.00	0.00

Bias includes components of proportional and constant errors. A proportional bias suggests an error of calibration whilst a constant bias suggests a blank error. Mixed errors will include significant components of both.

The Sy.x explained

Standard deviation

$$SD = \sqrt{\frac{\sum (\bar{y} - y)^2}{d.f.}}$$

where

\bar{y} = observed value

y = expected value

d.f. = degrees of freedom

Standard dev of residuals


$$Sy.x = \sqrt{\frac{\sum (\tilde{y} - y)^2}{d.f.}}$$

where

\tilde{y} = the value on the
line of best fit

Imprecision

Sodium (mmol/l)




■
 $r = 1.000$
 $IS = 0$
 $Sy.x = 0$


◆
 $r = 0.9812$
 $IS = 188$
 $Sy.x = 4.7$

Between batch precision

This Distribution L299



Previous Distributions



Current distribution

Interpretation of Accuracy Markers


Represents the point at which the line crosses the y axis. This is shown in the units of measurement of each analyte. In the example this is 12.04nmol/L. If deemed significant, this indicates a constant error.

$$y = mx + c$$


Represents the slope – in the example this is 0.97 which shows a 3% negative proportional bias. If deemed significant indicates a proportional error.

Accuracy Markers

This Distribution S245



Previous Distributions



Precision

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Sy.x = 6.13 nmol/L	Sy.x	4.46	27.15				
IS = 1	IS	2	53	0	0	0	0

Sy.x is the average deviation from the best fit line and is an index of scatter.

Precision Key

IS score	Interpretation
0 to 10	Good
11 to 150	Acceptable to Warning level
> 150	Unacceptable (including Curvilinear Data)

Accuracy


This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Systematic proportional error (calibration) -2.61%	Proportional (%)	2.11	3.08				
Systematic constant error (blank) 12.04 nmol/L	Constant (nmol/L)	-14.46	-5.24	0.00	0.00	0.00	0.00

Bias includes components of proportional and constant errors. A proportional bias suggests an error of calibration whilst a constant bias suggests a blank error. Mixed errors will include significant components of both.

Current slope and intercept
 $y = 0.97x + 12.04$

Previous slope and intercept values


ERRORS IN ACCURACY



ERRORS IN ACCURACY


1. NO ERROR

$y = 1x + 0$




ERRORS IN ACCURACY


2. SYSTEMATIC PROPORTIONAL (POSITIVE BIAS)




ERRORS IN ACCURACY 3. SYSTEMATIC PROPORTIONAL (NEGATIVE BIAS)



ERRORS IN ACCURACY 4. CONSTANT (POSITIVE BIAS)




ERRORS IN ACCURACY 5. CONSTANT (NEGATIVE BIAS)



ERRORS IN ACCURACY


6. MIXED



ERRORS IN ACCURACY

7. CURVILINEAR

—◆— curvilinear ($y=0.78x+1.3$)



Types of Errors

	Imprecision		Inaccuracy		
	Random	Curvilinear	Prop	Mixed	Constant
Slope	No	Yes/No	Yes	Yes	No
Intercept	No	Yes/No	No	Yes	Yes
$S_{y,x}$	Yes	Yes	No	No	No
r	Yes	Yes	No	No	No

Analyte detailed report

Scheme: Endocrine, Distribution Code: S245. Distribution Date: 24/10/17. Final Report Issued: 28/11/17							
Cortisol (nmol/L)	1	2	3	4	5	6	Analyte SDI
Reported Result	288.1	170.2	353.7	370.2	1064.0	72.0	
Method Corrected Result	288.10	170.20	353.70	370.20	1064.00	72.00	
Elecsys/E Module	Mean	306.62	184.24	381.54	396.64	1129.40	72.25
	SD	15.54	15.43	27.02	30.21	60.83	1.09
	Number	5	5	5	5	5	4
	Uncert.	6.950	6.903	12.083	13.512	27.204	0.545
Cobas E Module	Mean	306.62	184.24	381.54	396.64	1129.40	72.25
	SD	15.54	15.43	27.02	30.21	60.83	1.09
	Number	5	5	5	5	5	4
	Uncert.	6.950	6.903	12.083	13.512	27.204	0.545
Overall	Mean	282.70	163.62	347.38	359.41	1083.57	69.54
	SD	22.83	12.82	22.16	33.15	50.95	3.90
	Number	19	18	18	19	18	10
	Uncert.	5.237	3.022	5.224	7.606	12.009	1.233
Reference Values ID-LC-MS/MS	298.36	177.20	356.88	377.36	1065.45		
Ref. Value Uncertainty	6.100	3.600	7.300	7.700	21.800		
Non-scoring Reference Values							
WeQas SD	30.30	18.19	36.87	38.10	111.75	8.62	
SDI	0.18	0.36	0.17	0.28	-0.18	0.29	0.24

Please note: Linear regression uses CF corrected data.

Ref Values used for SDI and Sigma Scoring

Performance Criteria Target


Total Error

SDI is a measurement of your total error and will include both inaccuracy and imprecision.

This Distribution S245
Your average analyte SDI for the 6 samples is 0.24

Previous SDI

Distribution S245




--- Median —●— Your SDI — 97.5th


SDI Score

Analyte detailed report

This Distribution S245



Previous Distributions



Precision

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Sy.x = 6.13 nmol/L	Sy.x	4.46	27.15				
IS = 1	IS	2	53	0	0	0	0

Sy.x is the average deviation from the best fit line and is an index of scatter.

Precision Key

IS score	Interpretation
0 to 10	Good
11 to 150	Acceptable to Warning level
> 150	Unacceptable (including Curvilinear Data)

Accuracy

This Distribution S245	Previous Distributions	S244	S243	S242	S241	S240	S239
Systematic proportional error (calibration) -2.61%	Proportional (%)	2.11	3.08				
Systematic constant error (blank) 12.04 nmol/L	Constant (nmol/L)	- 14.46	- 5.24	0.00	0.00	0.00	0.00

Bias includes components of proportional and constant errors. A proportional bias suggests an error of calibration whilst a constant bias suggests a blank error. Mixed errors will include significant components of both.

Current slope and intercept
 $y = 0.97x + 12.04$

Previous slope and intercept values

Problem Solving Flow Chart

START HERE

IMPRECISION

INACCURACY

[1]
*Are you satisfied with
your imprecision values?
($Sy.x$, r)*

YES


[2]
*Are you satisfied with
slope and intercept?
(m , c)*

YES




Problem Solving Flow Chart

IMPRECISION




Problem Solving Flow Chart

START HERE



Problem Solving Flow Chart

INACCURACY




Problem Solving checklist

Analyte	SDI Score	Precision r , $Sy.x$, IS	Accuracy m,c	Previous dist.	Identify error	Possible Cause

Bias plot (1)

Sodium (mmol/l)



$$y = 0.9x + 9.6$$


$$r = 0.9995$$

$$IS = 5$$

$$Sy.x = 0.63$$

Bias plot (2)

Potassium (mmol/l)



Precision - blunder error

y = slope not calculated


r = 0.8826

IS = 1174

Sy.x = 1.10

Bias plot (3)

Chloride (mmol/l)



$$y = 1.01x - 4.01$$


$$r = 0.9995$$

$$IS = 5$$

$$Sy.x = 0.49$$

Bias plot (4)

Bicarbonate (mmol/l)



y = slope not calculated


r = 0.8484

IS = 1516

Sy.x = 2.41

Bias plot (5)

Urea (mmol/l)



$$y = 1.00x - 0.14$$


$$r = 1.000$$

$$IS = 0$$

$$Sy.x = 0.03$$

Bias plot (6)

Glucose (mmol/l)



$$y = 1.09x - 1.06$$


$$r = 0.9991$$

$$IS = 9$$

$$Sy.x = 0.40$$

Bias plot (7)

Calcium (mmol/l)



y = slope not calculated


r = 0.9810

IS = 190

Sy.x = 0.09

Bias plot (8)

ALP (IU/L)



$$y = 0.50 x + 1.16$$

$$r = 0.9999$$

$$IS = 1$$

$$S_{y.x} = 1.30$$